

NOTES:
 ① GAS OUTLETS TO BE MSC STANDARD BRAND OUTLET WITH CONNECTION TO MATCH EXISTING HOSPITAL EQUIPMENT.
 ② SERVICES REVERSED ON OPPOSITE HAND UNITS.
 ③ CEILING HEIGHT: 7'6" MIN TO 9'6" MAX. OTHER HEIGHTS ARE SUBJECT TO AN ADDITIONAL CHARGE.
 ④ CONTRACT YOUR SALES REP FOR DETAILS.
 ⑤ TWO SINGLE SIDED UNITS REQUIRED FOR BACK TO BACK INSTALLATIONS. TOTAL WALL THICKNESS WITH SINGLE LAYER OF 5/8" GYPSUM = 7.25".
 ⑥ APPROXIMATE JUNCTION BOX LOCATIONS.
 ⑦ MEDICAL GAS CONNECTIONS @ 85". EXACT HORIZONTAL SPACING TO BE DETERMINED IN SUBMITTALS.
 ⑧ INTERIOR STUD FRAMING 2-3/4" STUD WITH PRE-ATTACHED Z-BRACKET TO FIT A 3-5/8" METAL STUD WALL WITH 5/8" GYPSUM BOARD FINISH.
 ⑨ NO DRYWALL FASTENERS ARE TO BE USED IN THIS AREA.
 ⑩ PANEL LAMINATE COLOR TO BE DETERMINED. POWDER COATED SURFACES ARE TO MATCH FIBRICH #923 'SURF'.

LEGEND

REQ	SYM	DESCRIPTION
3	D	OXYGEN OUTLET
2	A	AIR OUTLET
4	V	VACUUM OUTLET
4	E2	DUPLEX RECEPTACLE, 8300 IVORY, 20A, 125V, H.G., NORMAL POWER
6	E4	DUPLEX RECEPTACLE, 8300 RED, 20A, 125V, H.G., EMER. POWER
1	NC	NURSE CALL PROVISION
1	CB	CODE BLUE PROVISION
1	MD	MONITOR OUTPUT PROVISION
1	MC	MONITOR CHANNEL ONLY
8	HC	HORIZONTAL ACCESSORY CHANNEL
QTY ACCESSORIES NOT SHOWN (PER UNIT)		

MODULAR 110 N. E. 38TH TERRACE
 SERVICES COMPANY OKLAHOMA CITY, OK. 73105

PROJECT

REVERSED

DUAL SILHOUETTE™
 SEMI-REVERSED
 MODEL 8600(SP)

SCALE: 1" = 1'

DATE: 08/22/08 DRAWN BY: DRAWING #: PB-196

