MODULAR SERVICES COMPANY 110 N. E. 38TH TERRACE OKLAHOMA CITY, OK. 73105

PROJECT

REVISED

CEILINGWORKS™
STRATUS
MODEL 5800

SCALE: =1'

DATE	DRAWN BY	DRAWING #
09/05/08		PB-163



NOTE:

(1) GAS OUTLETS TO BE MSC STANDARD BRAND OUTLET WITH CONNECTION TO MATCH EXISTING HOSPITAL EQUIPMENT

SEE STRATUS INSTALLATION INSTRUCTIONS FOR DETAILS ON MOUNTING COMPONENTS AND ROUGH-IN LOCATION REQUIREMENTS

-- REQ'D: --- AS SHOWN --- OPPOSITE







LEGEND

REQ	SYM	DESCRIPTION
2	O	OXYGEN OUTLET
2	A	AIR OUTLET
4	V	VACUUM OUTLET
4	E2	DUPLEX RECEPTACLE, 8300 IVORY, 20A, 125V, H.G., NORMAL POWER
11	E4	DUPLEX RECEPTACLE, 8300 RED, 20A, 125V, H.G., EMER. POWER
1	MO	MONITOR OUTPUT PROVISION
1	MC	MONITOR CHANNEL ONLY
1	BR	BRAKE RELEASE
4	VTT	VERTICAL EQUIPMENT MOUNTING TRACK, TRIPLE